

**Bid Document**

<b>Bid Details</b>	
Bid End Date/Time	28-12-2018 17:00:00
Bid Validity	90 (Days)
Ministry/State Name	Ministry Of Steel
Department Name	Na
Organisation Name	N/a
Office Name	Hyderabad
Total Quantity	2
Item Category	patient care beds for general use
MSE Exemption For Years Of Experience And Turnover	No

**Patient Care Beds For General Use ( 2 )**

**Technical Specifications**

[\\* As per GeM Category Specification](#)

<b>Specification</b>	<b>Specification Name</b>	<b>Values</b>	<b>Allowed Values</b>
DIMENSIONS OF MATTRESS PLATFORM & FRAME	<b>Length of mattress platform &amp; Overall length of bed</b>	>= 1900 mm & <=2000 mm	>= 1900 mm & <=2000 mm, >=1950 mm & <= 2050 mm, >=2000 mm & <=2100 mm, >= 2050 mm & <=2100 mm, >=2100 mm & <=2150 mm, >= 2150 mm & <= 2200 mm, >= 2200 mm & <= 2250 mm
	<b>Width of mattress platform</b>	900 mm to 920 mm	900 mm to 920 mm, >920 mm to 950 mm, >950 mm to 1000 mm, >1000 mm to 1050 mm
	<b>Material of mattress platform surface</b>	CRCA Sheet	CRCA Sheet, SS
	Height of mattress platform	678	*

	Mattress platform surface thickness	1.2 mm and above	*
	<b>Type of mattress platform</b>	With Holes	Perforated, With Holes
BED	Overall length of Bed(in mm):	1910	*
	FRAME section, dimensions, thickness and material	MS ERW rectangular tube with 50 mmx30 mm or more, gauge 1.2 mm or more.	*
	Frame width (in mm)	910	*
	Leg Shoe material	PVC	*
HEAD AND LEG BOARD	Height of Head board measured from mattress platform	355	*
	<b>Material of Leg board and head board</b>	MS tubular	SS laminated, MS tubular, Plastic
	Height of Leg board	355	*
	Pipe diameter of Head and Leg Boards	30 mm or more	*
	Thickness of pipe of Head and Leg board	1.2 mm or more	*
MISC	Maximum Load Capacity must be $\geq$ 150 Kg	Yes	*
	<b>Painting</b>	Powder coated	Powder coated
	Colour of Bed	white	*
	Provision of I.V. poles Holder	Yes	*
	Provision of Mosquito net pole Holder	No	*
Other details	Drawing available on Web.	No	*
	If Yes, Drawing No.	NA	*

\* Bold specifications are the golden parameters.

#### Consignees/Reporting Officer and Quantity

S.No.	Consignee/Reporting Officer	Address	Quantity	Delivery Days
1	Venkateswarlu Bandi	494553,NMDC LTD, BACHELI COMPLEX,BAILADILA, BACHELI PO, CHATTISGARH	2	30

**EMD Detail**

Required	No
----------	----

**ePBG Detail**

Required	No
----------	----

**Splitting**

Bid splitting not applied.

**Buyer Specific Additional Terms and Conditions**

- 1. Scope of supply (Bid price to include all cost components) :** Supply Installation Testing and Commissioning of Goods
- 2. Timely Servicing / rectification of defects during warrantee period:** After having been notified of the defects / service requirement during warrantee period, Seller has to complete the required Service / Rectification within time limit specified. If the Seller fails to complete service / rectification with defined time limit, a penalty of defined % of Unit Price of the product shall be charged as penalty for each week of delay from the seller. Seller can deposit the penalty with the Buyer directly else the Buyer shall have a right to recover all such penalty amount from the Performance Security (PBG).
  - (i) Time Limit for Service / Rectification of defects will be as defined in the SLA document for the service.
  - (ii) Penalty per week of delay as % of unit price of product will be as defined in the SLA document for the service.Seller to give compliance Yes / No while submitting bid.

[This Bid is also governed by the General Terms and Conditions](#)

**---Thank You---**